



Certificate of Analysis

PRODUCT: HYDROGEN PEROXIDE 30%

PRODUCT CODE: AJA260-2.5LPL

BATCH NO. : 396307 DATE MANUFACTURED: 6/10/2022

<u>Properties</u>	<u>Specification</u>	<u>Result</u>
Assay:	29.0 - 32.0% w/w	30.2%
Colour (APHA)	≤ 10	≤ 10
R.A.E.	≤ 0.002%	≤ 0.002%
Titrateable acid	≤ 0.0006 meq/g	≤ 0.0006 meq/g
Cl, Ca	≤ 0.00005% each	≤ 0.00005% each
H.M. (as Pb)	≤ 0.0001%	≤ 0.0001%
SO ₄ , NO ₃	≤ 0.0002% each	≤ 0.0002% each
NH ₄	≤ 0.0005%	≤ 0.0005%
Al	≤ 0.00003%	≤ 0.00003%
As	≤ 0.0003%	≤ 0.0003%
B, Mg, Ni	≤ 0.000005% each	≤ 0.000005% each
Na, Sn	≤ 0.001% each	≤ 0.001% each
Ti, Au, Cu, Pb, Zn	≤ 0.000002% each	≤ 0.000002% each
Cr, Mn	≤ 0.000001% each	≤ 0.000001% each
Fe	≤ 0.00001%	≤ 0.00001%
K	≤ 0.00002%	≤ 0.00002%
PO ₄	≤ 0.0005%	≤ 0.0005%

Retest Date: October 2025

7/10/2022

Signed:

Compliance Manager